



## **Postoperative stability following a triple pelvic osteotomy is affected by implant configuration: a finite element analysis**

Downloaded from: <https://research.chalmers.se>, 2024-04-19 16:46 UTC

Citation for the original published paper (version of record):

Hedelin, H., Brynskog, E., Larnert, P. et al (2022). Postoperative stability following a triple pelvic osteotomy is affected by implant configuration: a finite element analysis. *Journal of Orthopaedic Surgery and Research*, 17(1). <http://dx.doi.org/10.1186/s13018-022-03169-3>

N.B. When citing this work, cite the original published paper.

RESEARCH ARTICLE

Open Access



# Postoperative stability following a triple pelvic osteotomy is affected by implant configuration: a finite element analysis

Henrik Hedelin<sup>1,3\*</sup> , Erik Brynskog<sup>2</sup>, Per Larnert<sup>1,3</sup>, Johan Iraeus<sup>2</sup>, Tero Laine<sup>1,3</sup> and Kerstin Lagerstrand<sup>3,4</sup>

## Abstract

**Background:** The triple pelvic osteotomy is an established surgical method with multiple modifications regarding surgical technique and choice of implant. The stability of the osteotomy is affected by numerous factors, and among these, the three-dimensional implant configuration is a scientifically less explored aspect.

**Methods:** We used a finite element model of a hemi-pelvis with a standardized triple osteotomy to calculate relative flexibility for loads in all translational degrees of freedom for five different implant configurations. Two of the configurations used entry points only feasible when implant removal was not necessary.

**Results:** The stability of the osteotomy improved with an increased distance between the implants in the plane of the osteotomy as well as for a more perpendicular angle relative to the osteotomy plane. The implant configurations with more entry points available made this easier to adhere to.

**Conclusion:** The use of bioabsorbable implants may provide better opportunities for optimal implant constructs which can, to a certain degree, compensate for the lesser mechanical stiffness of bioabsorbable polymers as compared to metal implants.

**Keywords:** Osteotomy, Acetabulum, Fixation, Bioabsorbable screw, Finite element analysis

## Introduction

The triple pelvic osteotomy (TPO) is a re-directional osteotomy normally used in children over the age of 6–8 years old when the symphysis has become too stiff to allow a Salter osteotomy. The procedure is also an alternative to periacetabular osteotomies (PAO) in adults. Numerous modifications of the TPO have been proposed over the years with the Steel, Tönnis, and Carlioz modifications being the most well-known [1–3].

As in a Salter osteotomy, the TPO entails a cut of the ilium, and in addition also osteotomies of the ischial and

pubic bone. Since the entire acetabulum is mobilized, the TPO is inherently more unstable than most other pelvic osteotomies. Numerous different methods of fixation have been used for the iliac osteotomy, including K-wires, Steinmann pins, and screws and plates [4–6]. Recently, bioabsorbable screws have been presented as an alternative, negating the need for implant removal [7]. In clinical practice, many surgeons still use smooth K-wires for fixation [8], while some advocate threaded wires to reduce the risk of migration [9]. Fully threaded screws of varying diameter [10, 11] would be the choice by many surgeons in older children, while plate and screws solutions are rarely used.

The research regarding pelvic osteotomies has mainly been focusing on modifications of the osteotomies or type of implant used rather than how the positioning of the implants affects the stability of the osteotomy. The


\*Correspondence: [henrik.hedelin@vgregion.se](mailto:henrik.hedelin@vgregion.se)

<sup>1</sup> Department of Orthopaedics, Sahlgrenska University Hospital, Gothenburg, Sweden

Full list of author information is available at the end of the article



© The Author(s) 2022. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.



classic papers presenting the surgical method of a TPO or Salter osteotomy go into little detail regarding the exact placement of the implants. Salter simply states that “a stout Kirschner wire is inserted across the osteotomy site, through the graft and into the distal fragment posterior to the acetabulum in order to prevent any subsequent shift of the graft or of the distal fragment” [12]. Tönnis suggests the following fixation for his triple osteotomy: “The fragment is stabilized with four Kirschner wires 2.5 mm in diameter drilled in from the iliac crest. The wires are inserted in a fan-shaped pattern so that they diverge into the lateral, medial, anterior, and posterior parts of the acetabulum.” [13]. Carlöf, in his original article from 1982, also uses K-wires and mentions that he occasionally places a retrograde K-wire from an inferior/anterior point toward the proximal ilium, but he advises against this as routine practice due to difficulty in implant removal and that it may impair hip flexion [3].

Traditionally, the implants have been inserted from the iliac crest, in or close to the harvest site of the bone graft. The use of bioabsorbable screws, negating the need for implant removal, enables the surgeon to place the screws from any location that can be reached from the surgical approach [14]. Even though this could, based on basic orthopedic principles, provide a more advantageous biomechanical fixation, this has to our knowledge not been scientifically explored. One of the most elaborate studies on the subject is Yassir et al. who performed a cadaver-based study comparing the stability of three different screw configurations for a Tönnis triple osteotomy and a PAO [15]. In this study, the different

screw configurations in the iliac osteotomy did not, to a great extent, affect stability. The setup did not, however, include a graft and only examined weight-bearing loads and the results can therefore not be expected to be representative for a postoperative clinical setting.

To further evaluate the stability of an osteotomy for different implant configurations (*i.e.*, entry points, directions, and implant lengths), virtual models of mechanical systems can be used. The finite element (FE) method is a mathematical model to numerically solve partial differential equations that describe mechanical systems in fields such as structural analysis, heat transfer, and fluid flow [16]. When the FE method is used to model mechanical systems, the study geometry is discretized into smaller, simpler parts called finite elements, where each element is associated with a material model that describes its mechanical behavior. To complete the mathematical model, boundary conditions (*e.g.*, supports, connections with surrounding structures) and external loads (*e.g.*, force, pressure, prescribed motions) are applied (see Fig. 1). If the structure consists of several interacting parts, a contact algorithm can be applied preventing the parts to penetrate each other. FE analysis to simulate stress, strain, and stiffness within orthopedics has become increasingly popular [17, 18]. Vafaeian et al. performed an extensive review on FE analysis of acetabulum pressure related to surgery of developmental dysplasia of the hip [19]. Zou et al. used FE analysis to optimize the position of the acetabulum during a virtual PAO [20]. In addition, FE analysis has been used to study the influence of fixation parameters on biomechanical stability, *e.g.*, in

a proximal femoral osteotomy [21], sacroiliac joint fixations [22], or treatment of supracondylar humerus fractures [23]. Though virtual models can never completely replace clinical trials or ex vivo models like that of Yassir et al., they enable paired testing using the exact same “virtual” object and, as such, can be used to evaluate the stability of implants for different implant configurations.

In this study, we aim to evaluate how different implant configurations affect the stability of the iliac osteotomy in a TPO using finite element analysis with the purpose to explore if bioabsorbable screws can enable favorable biomechanical implant constructs that cannot be utilized using metal screws or K-wires.

## Materials and methods

A base FE model of an average shaped hemi-pelvis [24], consisting of the left ilium, the acetabulum, and a wedge-shaped graft (see description below) was created for simulations in Abaqus/Standard 2020.HF3 (Dassault systemes, Vélizy-Villacoublay, France). For information about the mathematical implementation to solve the FE equations, we refer to the Abaqus theory manual.<sup>1</sup> To illustrate and define loading directions for the current study, the pelvis was positioned in the standing anatomical position [25], see Fig. 1. The FE model consists of 46 000 tetrahedral solids (Abaqus element formulation C3D10), representing trabecular bone and 10 000 triangular shell elements (Abaqus element formulation STRI65), representing cortical bone, for an average element size of 3.2 mm for the solids and 2.9 mm for the shells. The cortical and trabecular bones were modeled using an isotropic linear elastic material model, with Young's modulus 17 000 MPa and 70 MPa, respectively [26]. For the cortical bone, a uniform thickness of 1 mm was assumed [26].

To simulate a triple osteotomy as described by Carioz [3], a virtual osteotomy was performed in the base geometry prior to meshing, superior to the acetabulum along with cuts in the ischial and pubic bones as illustrated in Fig. 1. The re-alignment of the acetabular fragment simulated a standardized correction with +15 deg Y+rotation and 5 deg X- rotation. The traditional harvest site at the superior anterior iliac spine and the more posterior alternative harvest site are visualized in Fig. 1. For the model, a standardized graft was harvested from the iliac wing (graft harvest site B, Fig. 1) and placed centrally in the osteotomy plane. This standardized graft was used for all models (even when entry points were from harvest site A). No elements

were removed from the harvest site assuming that the stability of the osteotomy and the acetabular fragment would be unaffected by the removal of these elements. The acetabular fragment was not moved in the X-axis, even though a displacement in X+ is mandated by some surgeons.

The implants were modeled using beam elements with a circular cross section of 4.5 mm in diameter and using an isotropic linear elastic material model with a Young's modulus of 5 000 MPa.<sup>2</sup> All nodes that belonged to the ilium, the graft, and the acetabulum, enclosed by the cylinder representing the 4.5-mm-diameter screw, were fully constrained (Abaqus \*TIE) to the implant beam elements. These nodes are highlighted as white dots in Fig. 1. This simulates a fully threaded screw, and as such, no beam pretension was applied. This means that the contact stress was zero at the start of each simulation.

To model the boundary conditions, the ilium was constrained in all translational and rotational DOF at the SI joint surface, see Fig. 1, while the distal fragment was considered un-constrained. Contacts (Abaqus \*CONTACT PAIR, SURFACE\_TO\_SURFACE, with friction coefficient 0.2), as suggested<sup>3</sup> to be used together with element type C3D10, were defined between ilium and the graft as well as the acetabulum and the graft.

Pelvis loading was applied at the center of the femoral head, for one direction (X/Y/Z translation or X/Y/Z rotation) at a time, using prescribed unit displacements. The load, 1 mm translation or 1 degree rotation, was distributed over the lunate surface of acetabulum (using Abaqus kinematic \*COUPLING, degrees of freedom (DOF) 1–6).

Finally, to be easily comparable to the work by Yassir et al., stability of the virtual osteotomy was evaluated in terms of mechanical flexibility<sup>4</sup> (i.e., displacement for a given load measured in [mm/N]), where higher stability corresponds to lower flexibility. For each implant configuration, the relative flexibility from translational loads was evaluated in all DOFs except for Z-, describing a clinically unlikely distracting force on the osteotomy. Relative flexibility from rotational loads was calculated for all DOFs. Relative, in this context, means that the average flexibility was computed for all implant configurations, and then, each configuration was normalized by this average. This was performed

<sup>2</sup> Reported characteristics of a 4.5-mm 85L/15G Poly lactic-co-glycolic acid (PLGA) cannulated screw, previously used for pelvic osteotomies, by the manufacturer of Activa screw (Bioretec, Tampere, Finland).

<sup>3</sup> ABAQUS (2016) Getting Started with Abaqus/CAE. Version 2016, Dassault Systemes Simulia, Inc.

<sup>4</sup> Flexibility is the inverse of stiffness, where stiffness is defined as resistance to deformation in response to applied force and measured in [N/mm].

<sup>1</sup> ABAQUS (2016) Theory Manual. Version 2016, Dassault Systemes Simulia, Inc.

separately for each direction and DOF. Since the load was applied at the center of acetabulum, the lever arm caused any translation in X and Y to be coupled by rotation. For example, a rotation around Y+ will correspond to a translation in X-. A rotation around the Z-axis, however, does not couple with a translation since no translation include a lever arm on this axis.

For the present simulation, five implant placement configurations, corresponding to five different FE models, were chosen to represent different clinically feasible fixation solutions of the osteotomy (Table 1). In configurations A, B, and C, the entry points were chosen to enable later implant removal, using the classic harvest site of the graft from the iliac crest (harvest site A, Fig. 1). D and E were configured to include surgically feasible entry points that did not need to be easily accessible for later extraction. These entry points included the inner and outer aspects of the iliac wing, as well as entry points from a more proximal harvest site (Fig. 1, harvest site B and Table 1, models D and E). The principles behind the implant configurations for respective model are presented in Table 1. For all chosen placement configurations, three principles were adhered to: (I) Two out of three screws should pass through the triangular graft, (II) the screws should, if not otherwise stated, not end superior to the acetabulum, and (III) the distance between the implants in the graft plane should be maximized, considering surgical constraints and I and II.

## Results

Figure 2 shows the resulting flexibility for different implant placements. The most relevant results from a clinical point of view (X-, Y-, Z+) are highlighted with dashed boxes. The values for X+ (lateralization of the distal fragment) and Y+ (anterior displacement of the distal fragment) are presented for completeness but usually do not present a problem in clinical practice. Tests for Z- were, as mentioned before, omitted due to distraction never being a relevant factor in this case. The flexibility for rotational loads behaved, like expected, in correspondence with the translational loads, and the results for rotation in X+ mirrored the translation in Y+ almost exactly. The complete results for rotational loads are available in Additional file 1: Appendix 1.

Averaging the relative translational flexibility in Fig. 2, over the three relevant directions (Z+, X-, and Y-), results in average flexibilities from 0.82 for model E to 1.35 for model B (model A = 1.03, Model C and D = 0.90), meaning that model E is on average the least flexible and model B the most flexible. To add some granularity, model E showed the most consistent low flexibility of all models with 18% less flexibility compared to the average flexibility. Model A shows lower flexibility than model

B in the clinically relevant X-, Y-directions (however, both have 3% and 35% higher flexibility compared to the average) and roughly equivalent in Z+. Model C, with a screw placed from the distal fragment, was equivalent or less flexible in all degrees of freedom compared to A and B and also performed on par with D and E in all dimensions except Y- where C showed more flexibility. It can also be noted that D, with a flexibility comparable to C and E in most dimensions, has a markedly higher flexibility in X-. The Z+ movement, which largely corresponds to a closing of the osteotomy, exhibited similar flexibility for all implant positions.

## Discussion

The FE method has been proven to be useful for postoperative modeling of different orthopedic applications and has been used to evaluate different fixation methods for PAO [20, 27, 28]. Despite TPO being a mainstay of advanced surgical treatment of multiple conditions, FE analysis has, to our knowledge, not previously been applied to evaluate the stability of different implant constructs.











In this study, we present a FE analysis of how screw placements affect the stability of a standardized TPO. Model E, with the most entry points available, clearly had the best results from a stability point of view. Model C, with a distally placed screw, also showed high stability, but placing a screw from the acetabular fragment has drawbacks. Firstly, extraction of a screw (or pin) from this entry point can be challenging, secondly the antero-lateral corner superior to the acetabulum is small and can potentially break, and thirdly implants in this location may also interfere with hip abduction. In the other models with all screws placed from the superior fragment (A, B and D), the option to place the screws from more entry points (D) outperformed the models with less entry points available (A and B).


Adamczyk et al. used an in vitro model to compare the stability of bioabsorbable (PLLA<sup>5</sup>) and metal screws after a TPO [29] in weight-bearing conditions. No significant difference was found at physiological loads, when comparing the two materials in both the standing and spica cast positions. This was likely, according to the authors, due to the inherent stability of the remaining intact pelvic ring, as compression of the osteotomy will cause bony contact that carry the load regardless of implant material. This result is in line with our findings for Z+ loading, where implant position had a limited effect on stability when compressing the osteotomy. Hence, it can be

<sup>5</sup> Poly-(L)-lactic acid. A polymer comparable, but not identical, to 85L/15G PLGA screws used in this article.



**Table 1** Description of implant placements for the five FE models. The model is of the left ilium and acetabulum. The frontal view shows the pelvis and implants in a posterior view, and the lateral in a medial view. The harvest site is indicated with dashed lines

Model	Entry points used	Principle behind screw placement
<b>A</b> <div> <div>Frontal</div>  <div>Lateral</div>  </div>	<ul style="list-style-type: none"> <li>• Implant accessible for removal</li> <li>• Harvest site A</li> </ul>	<ul style="list-style-type: none"> <li>• Three implants placed from the graft site at the iliac wing</li> <li>• Maximum spread from the entry point</li> <li>• The most anterior screw ends superior to the acetabulum</li> </ul>
<b>B</b> <div> <div>Frontal</div>  <div>Lateral</div>  </div>	<ul style="list-style-type: none"> <li>• Implant accessible for removal</li> <li>• Harvest site A</li> </ul>	Similar to A but sacrificing optimal spread to avoid any screw ending superior to the acetabulum.
<b>C</b> <div> <div>Frontal</div>  <div>Lateral</div>  </div>	<ul style="list-style-type: none"> <li>• Implant accessible for removal</li> <li>• Harvest site A</li> </ul>	Similar to B but the third implant placed in a retrograde fashion from the lateral corner of the acetabulum
<b>D</b> <div> <div>Frontal</div>  <div>Lateral</div>  </div>	<ul style="list-style-type: none"> <li>• Implant not easily accessible for removal</li> <li>• Harvest site B</li> </ul>	<ul style="list-style-type: none"> <li>• Two implants placed from the alternative graft site as perpendicular as possible</li> <li>• One implant placed from the antero-medial aspect of the ilium.</li> </ul>
<b>E</b> <div> <div>Frontal</div>  <div>Lateral</div>  </div>	<ul style="list-style-type: none"> <li>• Implant not easily accessible for removal</li> <li>• Harvest site B</li> </ul>	<ul style="list-style-type: none"> <li>• Aiming for maximal spread of implants using any superior entry point available from any harvest site or part of the iliac wing</li> </ul>



suspected that the osteotomy and graft position are more important than the strictly mechanical properties of the implant when studying a weight-bearing scenario. For loading directions where the implants are mainly subjected to bending forces, e.g., in the X- and Y-directions, the resulting stability will be a trade-off between strong material properties and optimal placement. Bioabsorbable screws enable a superior implant placement which should be considered along with other benefits, like not needing extraction and less risk of implant migration, when making a judgment for the best procedure. Our present findings show, in line with two basic orthopedic principles, that implants should be spread out in all dimensions over the osteotomy and that implant placements more perpendicular to an osteotomy improve stability. Both principles are easier to adhere to when more entry points are available. Bioabsorbable screws are not as strong as an equivalent metal screw, but the benefit of more placement options may, to a certain degree, compensate this limitation. A TPO can, however, also be performed on adults, and in these cases, extraction of metal screws may not be necessary and the benefit and indication of bioabsorbable implants thus decrease.

In the present study, we have not used physiological loads but rather focused on a generic prescribed displacement of 1 mm or 1 degree in a given direction. The displacements for different fixation methods are in the same order of magnitude as found by Yassir et al. [11] (<2 mm displacement and <3 deg rotation), supporting the load used in the present study.

A load analysis based on everyday motions or standing loads might have modified the results [30], but this was out of the scope of the study. Since the TPO is normally followed either by a hip spica, an orthosis or non-weight-bearing in a wheelchair, physiological loads in a standing position are of less relevance until bony healing has occurred (6–8 weeks postoperatively). The present study, therefore, focused on these, crucial, first weeks postoperatively rather than stability during a later stage.

A validation of the FE model by comparing the model predictions to a physical test was not feasible in the current study, since no physical experiment measuring stiffness in all DOFs exists in the literature. The only relevant study with a similar test setup, performed on cadaver specimens, is Yassir et al. [15] enabling a plausibility check of the model predictions in vertical loading (Z+ in the current study). In the study by Yassir et al., the authors report a 0.7 mm deformation for a screw configuration similar to present Model C, and a 1.2 mm deformation for a screw configuration similar to present Model A/B with vertical loading (Z+ in the current study), for an external load of 450 N (equivalent to 225 N per hip joint). The Z+ stiffness in the current study ranged from 626 to 748 N/mm (see Additional file 1: Appendix) for Models A–C, corresponding to a 0.30–0.36 mm deformation. This is slightly smaller to the deformations measured by Yasser et al., but it should be pointed out that the FE model used in this study represents a “perfect” osteotomy, with all bony

parts perfectly cut and aligned. This is not the case in reality, where any imperfections will result in larger deformation. In addition, Yasser et al. point out the risk of “increased osteotomy fragment motion secondary to screw loosening in the later trials,” regarding the repeated testing. In the light of this, the model predictions seem to be plausible, at least for vertical loading in the Z+ direction.

In the present model, the influence of muscle forces and soft tissues was also discarded. Muscles and soft tissues can provide stability to the pelvis, while on the other hand, muscle forces can cause greater loading [31]. Due to the partial immobilization of the patient following a triple pelvic osteotomy, the variation in loading from muscles forces is considered to be of less importance.

All displacements are not of equal clinical relevance, and a distracting force on the osteotomy is, for example, irrelevant. In clinical practice, it is widely known that the movements normally are associated with instability and a collapse of an osteotomy is firstly a “closing” of the osteotomy (mostly Z+ movement in our model) and/or a postero-medial dislocation of the distal fragment (X– in our model). A rotation around the X-axis affects the ante/retroversion of the acetabulum with both being undesirable but an increased retroversion even less so. A displacement in both X– and Y– negates the correction achieved and means that the distal fragment may lose the hinge of cortical bone in the infero-medial corner of the osteotomy.

Due to the complexity of the three-dimensional shape of the pelvis, the theoretically possible entry points and subsequent angles and lengths of an implant are innumerable and the modeling had to be restricted to study only a few clinically feasible solutions. Our study does not present reference values but intend to serve as a guide for surgical planning of TPO by outlining factors affecting stability and support. Moreover, our study was performed with a singular type of implant, corresponding to a 4.5-mm resorbable 85L/15G PLGA screw. Compared to commonly used 2-mm-steel K-wires, the 4.5-mm bioabsorbable screw has 60% less bending stiffness. Thus, for loading directions where the implants are mainly subjected to bending forces, e.g., in the X- and Y-directions, the joint overall flexibility should be comparable. For loading directions where the implants are mainly subjected to axial loads, (Z–), a 2-mm K-wire is almost 10 times stiffer than a 4.5-mm bioabsorbable screw, but this is of less clinical relevance. Also, lacking threads, the K-wire cannot fully utilize its axial stiffness, so the pull-out strength of a 4.5-mm bioabsorbable screw remains superior. Depending on the specific type of implant, the absolute magnitude of the results will vary to some extent, but the relative stability/flexibility

between different implant configurations is expected to be consistent.

Although the study demonstrated the comparison of different fixation systems, the modeling had its limitations. Similar to other established finite element models of the pelvis [28], the implant screws were simplified as homogeneous with isotropic materials. Furthermore, a constant cortical thickness and homogeneous and isotropic material properties were assigned to the pelvis bone. This is a simplification compared to some other models in the literature that include a non-homogeneous distribution of trabecular bone properties [28, 32, 33]. However, since the aim of this study was to compare the biomechanical performance of different implant configurations in identical conditions, and not to study the stress–strain state of the bone or the contact stress of the osteotomy, this simplification was justified. This is further motivated by analyzing differences of relative flexibility, as a small error in structural stiffness would be treated equally for all configurations by the normalization of the results.

Using a similar motivation, we have not performed any mesh converges study. Since flexibility (or stiffness) relates to the displacement while stress/strain relates to the derivative of the displacement, structural flexibility will converge much faster than stress/strain. Therefore, mesh convergence studies are of greater importance for stress or strain analyses. A relevant example is presented in the ABAQUS (2016) user guideline, where even a very coarse mesh consisting of just 14 elements gives a stiffness prediction within 3% of the accurate result, while the stress at the critical area is not converged even using the finest mesh consisting of 1800 elements.

Moreover, the individual shape of the pelvis may vary greatly, affected by sex, age, and as shaped by the underlying condition necessitating surgery. The shape of the pelvis should not, however, greatly affect the relative stability between the screw configurations. Nevertheless, additional studies are warranted to evaluate the applicability of our findings in clinical practice. While virtual models cannot completely replace experimental testing, they provide a valuable tool that is rapidly evolving.

## Conclusion

In a FE model of a TPO, different implant placement configurations had a substantial impact on the stability of the osteotomy. As predicted, a greater spread of the implants over the osteotomy and configurations more perpendicular to the osteotomy improved the stability. The findings suggest that bioabsorbable screws in TPO, enabling a higher degree of freedom when selecting



entry points, provide superior preoperative stability as compared to TPO with only traditional entry points.

### Abbreviations

FE: Finite element; MPa: Mega Pascal; PAO: Periacetabular osteotomy; PLGA: Poly-lactic-co-glycolic acid; PLLA: Poly-(L)-lactic acid; SO: Salter osteotomy; TPO: Triple pelvic osteotomy.

## Supplementary Information

The online version contains supplementary material available at <https://doi.org/10.1186/s13018-022-03169-3>.

**Additional file 1: Appendix 1.** Tabulation of all results from the FE simulations.

### Acknowledgements

Not applicable

### Authors' contributions

HH was main author and contributed to study design. EB was responsible for FE analysis and interpretation. PL was co-author and orthopedic consultant and involved in review of manuscript. JJ was senior consultant regarding FE analysis and interpretation and contributed to review of manuscript. TL was consultant initiating study and study design and involved in managing of figures and tables. KL was senior technical consultant and contributed to review of manuscript. All authors read and approved the final manuscript.

### Funding

Open access funding provided by University of Gothenburg. This work was supported by a grant from Göteborgs Läkaresällskap (Gothenburg Medical Society) GLS-934779.

### Availability of data and materials

All data generated or analyzed during this study is included in the published article, including Additional file 1: Appendix.

### Declarations

#### Ethics approval and consent to participate

Ethical approval for the study was obtained (Dnr: 2020–02872).

#### Consent for publication

Not applicable.

#### Competing interests

The authors declare that they have no competing interests.

#### Author details

<sup>1</sup>Department of Orthopaedics, Sahlgrenska University Hospital, Gothenburg, Sweden. <sup>2</sup>Department of Mechanics and Maritime Sciences, Chalmers University of Technology, Gothenburg, Sweden. <sup>3</sup>Institute of Clinical Sciences, Sahlgrenska Academy, University of Gothenburg, Gothenburg, Sweden. <sup>4</sup>Medical Physics and Biomedicine, Sahlgrenska University Hospital, Gothenburg, Sweden.

Received: 24 January 2022 Accepted: 9 May 2022

Published online: 15 May 2022

### References

- Steel HH. Triple osteotomy of the innominate bone. *J Bone Jt Surg Am*. 1973;55(2):343–50.
- Tonnis D, Behrens K, Tscharani F. A modified technique of the triple pelvic osteotomy: early results. *J Pediatr Orthop*. 1981;1(3):241–9.
- Carlioz H, Khouri N, Hulin P. Triple juxtacotyloid osteotomy. *Rev Chir Orthop Reparatrice Appar Mot*. 1982;68(7):497–501.
- Kamaci S, Kaymakoğlu M, Ramazanov R, Aksoy MC, Yılmaz G. Preliminary results of rigid fixation (locking plate/screw) after triple pelvic osteotomy without a hip spica cast: a modified fixation method in 21 patients. *Jt Dis Relat Surg*. 2021;32(2):454–60.
- Rosello O, Solla F, Oborocianu I, Chau E, ElHayek T, Clement JL, et al. Advanced containment methods for Legg–Calve–Perthes disease: triple pelvic osteotomy versus Chiari osteotomy. *Hip Int J Clin Exp Res Hip Pathol Ther*. 2018;28(3):297–301.
- Stepanovich M, Upasani VV, Bomar JD, Wenger DR. Advanced containment with triple innominate osteotomy in Legg–Calve–Perthes disease: a viable option even in severe cases. *J Pediatr Orthop*. 2017;37(8):563–9.
- Hedelin H, Larnert P, Antonsson P, Lagerstrand K, Brisby H, Hebelka H, et al. Stability in pelvic triple osteotomies in children using resorbable PLGA screws for fixation. *J Pediatr Orthop*. 2021.
- Hailer NP, Soykaner L, Ackermann H, Rittmeister M. Triple osteotomy of the pelvis for acetabular dysplasia: age at operation and the incidence of nonunions and other complications influence outcome. *J Bone Jt Surg Br*. 2005;87(12):1622–6.
- Sales de Gauzy J. Pelvic reorientation osteotomies and acetabuloplasties in children. Surgical technique. *Orthop Traumatol Surg Res*. 2010;96(7):793–9.
- Lipton GE, Bowen JR. A new modified technique of triple osteotomy of the innominate bone for acetabular dysplasia. *Clin Orthop Relat Res*. 2005;434:78–85.
- Wenger DR, Pandya NK. Advanced containment methods for the treatment of Perthes disease: Salter plus varus osteotomy and triple pelvic osteotomy. *J Pediatr Orthop*. 2011;31(2 Suppl):S198–205.
- Salter RB. Innominate osteotomy in the treatment of congenital dislocation and subluxation of the hip. *Bone Jt J*. 1961;43(3):518–39.
- Tönnis PDD, editor. Congenital dysplasia and dislocation of the hip in children and adults. Berlin: Springer; 1987.
- Hedelin H, Larnert P, Hebelka H, Brisby H, Lagerstrand K, Laine T. Innominate Salter osteotomy using resorbable screws: a retrospective case series and presentation of a new concept for fixation. *J Child Orthop*. 2019;13(3):310–7.
- Yassir W, Mahar A, Aminian A, Newton P, Wenger D. A comparison of the fixation stability of multiple screw constructs for two types of pelvic osteotomies. *J Pediatr Orthop*. 2005;25(1):14–7.
- Mediouni M, Schlatterer DR, Khoury A, Von Bergen T, Shetty SH, Arora M, et al. Optimal parameters to avoid thermal necrosis during bone drilling: a finite element analysis. *J Orthop Res*. 2017;35(11):2386–91.
- Dong E, Wang L, Iqbal T, Li D, Liu Y, He J, et al. Finite element analysis of the pelvis after customized prosthesis reconstruction. *J Bionic Eng*. 2018;15(3):443–51.
- Xu M, Qu W, Wang Y, Zhong L, Zhu Z, Li W, et al. Theoretical implications of periacetabular osteotomy in various dysplastic acetabular cartilage defects as determined by finite element analysis. *Med Sci Monit*. 2016;22:5124–30.
- Vafaeian B, Zonoobi D, Mabee M, Hareendranathan AR, El-Rich M, Adeeb S, et al. Finite element analysis of mechanical behavior of human dysplastic hip joints: a systematic review. *Osteoarthritis Cartil*. 2017;25(4):438–47.
- Zou Z, Chavez-Arreola A, Mandal P, Board TN, Alonso-Rasgado T. Optimization of the position of the acetabulum in a ganz periacetabular osteotomy by finite element analysis. *J Orthop Res*. 2013;31(3):472–9.
- Tai CL, Chen WP, Chen HH, Lin CY, Lee MS. Biomechanical optimization of different fixation modes for a proximal femoral L-osteotomy. *BMC Musculoskelet Disord*. 2009;10:112.
- Lee CH, Hsu CC, Huang PY. Biomechanical study of different fixation techniques for the treatment of sacroiliac joint injuries using finite element analyses and biomechanical tests. *Comput Biol Med*. 2017;87:250–7.
- Durusoy S, Oner K, Ozer A, Sevinc HF. The effect of the angle between fracture line and Kirschner wires on stability in supracondylar humerus fractures treated with Kirschner wire fixation: a finite element analysis. *Jt Dis Relat Surg*. 2021;32(1):75–84.
- Brynskog E, Iraeus J, Reed MP, Davidsson J. Predicting pelvis geometry using a morphometric model with overall anthropometric variables. *J Biomech*. 2021;126:110633.

25. Reynolds HM, Snow CC, Young JW. Spatial geometry of the human pelvis. In: Administration FA (ed) Oklahoma City; 1982.
26. Dalstra M, Huiskes R, van Erning L. Development and validation of a three-dimensional finite element model of the pelvic bone. *J Biomech Eng.* 1995;117(3):272–8.
27. Babis GC, Trousdale RT, Jenkyn TR, Kaufman K. Comparison of two methods of screw fixation in periacetabular osteotomy. *Clin Orthop Relat Res.* 2002;403:221–7.
28. Bakhtiarinejad M, Farvardin A, Murphy RJ, Grupp RB, Tis JE, Sponseller PD, et al. Comparative biomechanical study of screw fixation techniques in periacetabular osteotomy. *Biomechanics.* 2021;1(1):131–44.
29. Adamczyk MJ, Odell T, Oka R, Mahar AT, Pring ME, Lalonde FD, et al. Biomechanical stability of bioabsorbable screws for fixation of acetabular osteotomies. *J Pediatr Orthop.* 2007;27(3):314–8.
30. Agarwal Y, Doebele S, Windolf M, Shiozawa T, Gueorguiev B, Stuby FM. Two-leg alternate loading model—a different approach to biomechanical investigations of fixation methods of the injured pelvic ring with focus on the pubic symphysis. *J Biomech.* 2014;47(2):380–6.
31. Sverdlova NS, Witzel U. Principles of determination and verification of muscle forces in the human musculoskeletal system: Muscle forces to minimise bending stress. *J Biomech.* 2010;43(3):387–96.
32. Anderson AE, Peters CL, Tuttle BD, Weiss JA. Subject-specific finite element model of the pelvis: development, validation and sensitivity studies. *J Biomech Eng.* 2005;127(3):364–73.
33. Zhang Q-H, Wang J-Y, Lupton C, Heaton-Adegbile P, Guo Z-X, Liu Q, et al. A subject-specific pelvic bone model and its application to cemented acetabular replacements. *J Biomech.* 2010;43(14):2722–7.

## Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

**Ready to submit your research? Choose BMC and benefit from:**

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

**At BMC, research is always in progress.**

Learn more [biomedcentral.com/submissions](https://biomedcentral.com/submissions)

